

## New England Biolabs Certificate of Analysis

Product Name: *HiScribe™ T7 ARCA mRNA Kit (with Tailing)*  
 Catalog Number: *E2060S*  
 Lot Number: *10015052*  
 Expiration Date: *01/2020*  
 Storage Temperature: *-20°C*  
 Specification Version: *PS-E2060S v1.0*

HiScribe™ T7 ARCA mRNA Kit (with Tailing) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N2053AVIAL	2X ARCA/NTP Mix	10011908	Pass
N0247AVIAL	CLuc Control Template	10011907	Pass
M0444AVIAL	E. coli Poly(A) Polymerase	10012624	Pass
M0303AAVIAL	DNase I (RNase-free)	10012626	Pass
M0255AAVIAL	T7 RNA Polymerase Mix	0161803	Pass
B2051AVIAL	LiCl Solution	10012625	Pass
B0276SVIAL	Poly(A) Polymerase Reaction Buffer	0131804	Pass

Assay Name/Specification	Lot # 10015052
<p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in HiScribe™ T7 ARCA mRNA Kit (with tailing) and meet the designated specifications.</p>	Pass
<p><b>Functional Testing (RNA Tailing, Length)</b>            A 50 µl reaction in Poly(A) Polymerase Reaction Buffer containing a minimum of 15 µg of HiScribe™ RNA product and 5 µl of Poly(A) Polymerase incubated for 30 minutes at 37°C results in a visible shift of the 1.6 kb RNA band as determined by agarose gel electrophoresis.</p>	Pass
<p><b>Functional Testing (RNA Synthesis, Yield)</b>            A 20 µl reaction in the presence of 1 X ARCA/NTP Mix containing 1 µg CLuc Control Template with 2 µl of the T7 RNA Polymerase Mix incubated for 30 minutes at 37°C results in a minimum of 15 µg of a 1.6 kb RNA as determined by OD 260 and agarose gel electrophoresis.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.



---

Bhairavi Jani  
Production Scientist  
27 Jul 2018



---

Michael Tonello  
Packaging Quality Control Inspector  
27 Jul 2018