

## New England Biolabs Certificate of Analysis

**Product Name:** NEBuffer™ 2.1  
**Catalog #:** B7202S  
**Concentration:** 10X Concentrate  
**Lot #:** 0241705  
**Assay Date:** 05/2017  
**Expiration Date:** 05/2020  
**Storage Temp:** -20°C  
**Composition (1X):** 50 mM NaCl, 10 mM Tris-HCl, 10 mM MgCl<sub>2</sub>, 100 µg/ml BSA, (pH 7.9 @ 25°C)  
**Specification Version:** PS-B7202S v1.0  
**Effective Date:** 31 Jan 2018

Assay Name/Specification (minimum release criteria)	Lot #0241705
<b>Conductivity (buffers/solutions)</b> - The conductivity of 10X NEBuffer 2.1 is between 55 and 62 mS at 25°C.	<b>Pass</b>
<b>Endonuclease Activity (Nicking, Buffer)</b> - A 50 µl reaction in 1X NEBuffer 2.1 containing 1 µg of supercoiled PhiX174 DNA incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	<b>Pass</b>
<b>Functional Testing (Restriction Digest, Buffer)</b> - A 50 µl reaction in 1X NEBuffer 2.1 containing 1 µg of Lambda DNA and 1 unit of HindIII incubated for 1 hour at 37°C results in complete digestion of the substrate DNA as determined by agarose gel electrophoresis.	<b>Pass</b>
<b>Functional Testing (Restriction Digest, Buffer)</b> - A 50 µl reaction in 1X NEBuffer 2.1 containing 1 µg of Lambda DNA and 1 unit of SphI incubated for 1 hour at 37°C results in complete digestion of the substrate DNA as determined by agarose gel electrophoresis.	<b>Pass</b>
<b>Non-Specific DNase Activity (16 hour, Buffer)</b> - A 50 µl reaction in 1X NEBuffer 2.1 containing 1 µg of PhiX174-HaeIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	<b>Pass</b>
<b>pH (buffers/solutions)</b> - The pH of 10X NEBuffer 2.1 is between pH 7.8 and 8.0 at 25°C.	<b>Pass</b>
<b>RNase Activity (Buffer)</b> - A 10 µl reaction in 1X NEBuffer 2.1 containing 40 ng of a 300 base single-stranded RNA is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by fluorescent detection.	<b>Pass</b>



---

## New England Biolabs Certificate of Analysis

\* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.



---

Authorized by  
Derek Robinson  
31 Jan 2018



---

Inspected by  
Tony Spear-Alfonso  
30 May 2017

