

## New England Biolabs Certificate of Analysis


Product Name: *phi29-XT RCA Kit*  
 Catalog Number: *E1603L*  
 Packaging Lot Number: *10183353*  
 Expiration Date: *02/2025*  
 Storage Temperature: *-20°C*  
 Specification Version: *PS-E1603S/L v1.0*

phi29-XT RCA Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S1884LVIAL	Exonuclease-Resistant Random Primers	10182219	Pass
N0447XVIAL	Deoxynucleotide (dNTP) Solution Mix	10182218	Pass
M0572LVIAL	phi29-XT DNA Polymerase	10182217	Pass
B0572SVIAL	phi29-XT Reaction Buffer	10182220	Pass

Assay Name/Specification	Lot # 10183353
<p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in phi29-XT RCA Kit and meet the designated specifications.</p> <p><b>Functional Testing (RCA DNA Amplification)</b>            A 20 µl rolling circle amplification (RCA) reaction in 1X phi29-XT Reaction Buffer in the presence of 1X phi29-XT DNA Polymerase and 50 µM Exonuclease-Resistant Random Primers generates at least 5 µg DNA from 0.1 ng plasmid DNA after incubation for 2 hours at 42°C.</p>	<p>Pass</p> <p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.

  
Lea Antonopoulos  
Production Scientist  
27 Mar 2023

  
Michael Tonello  
Packaging Quality Control Inspector  
04 Apr 2023