

New England Biolabs Certificate of Analysis

Product Name: Luna® Universal qPCR Master Mix
Catalog Number: M3003X
Concentration: 2 X Concentrate
Lot Number: 10029537
Expiration Date: 09/2020
Storage Temperature: -20°C
Specification Version: PS-M3003S/L/G/X/E v1.0
Composition (1X): Proprietary


| Luna® Universal qPCR Master Mix Component List | | | |
|--|---------------------------------|------------|----------------------|
| NEB Part Number | Component Description | Lot Number | Individual QC Result |
| M3003L | Luna® Universal qPCR Master Mix | 10025104 | Pass |

| Assay Name/Specification | Lot # 10029537 |
|---|----------------|
| <p>Functional Testing (qPCR) Luna® Universal qPCR Master Mix is functionally tested in qPCR with human cDNA template, resulting in a standard curve with a calculated qPCR efficiency of 90-110%, and a dynamic range of 5 orders of magnitude.</p> | Pass |
| <p>Non-Specific DNase Activity (16 hour, Master Mix) A 50 µl reaction in 1X Luna® Universal qPCR Master Mix containing 1 µg of T3 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p> | Pass |
| <p>qPCR DNA Contamination (E. coli Genomic) A minimum of 1 µl of Luna® Universal qPCR Master Mix is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.</p> | Pass |
| <p>RNase Activity Assay (4 Hour Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of Luna® Universal qPCR Master Mix is incubated at 37°C. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.</p> | Pass |

This product has been tested and shown to be in compliance with all specifications.



Christie Vazquez
Production Scientist
21 Nov 2018



Michael Tonello
Packaging Quality Control Inspector
21 Nov 2018