

New England Biolabs Certificate of Analysis

Product Name: Luna® Probe One-Step RT-qPCR 4X Mix with UDG (No ROX)
Catalog Number: M3029L
Concentration: 4 X Concentrate
Packaging Lot Number: 10176923
Expiration Date: 10/2024
Storage Temperature: -20°C
Specification Version: PS-M3029S/L/E v2.0
Composition (1X): Proprietary


Luna® Probe One-Step RT-qPCR 4X Mix with UDG (No ROX) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M3029LVIAL	Luna Probe One-Step RT-qPCR 4X Mix with UDG (No ROX)	10176924	Pass
B1502AVIAL	Nuclease-free Water	10154305	Pass

Assay Name/Specification	Lot # 10176923
Functional Testing (One-Step RT-qPCR) Luna® Probe One-Step RT-qPCR 4X Mix with UDG (No ROX) is functionally tested in one-step RT-qPCR with human RNA template, resulting in a standard curve with a calculated qPCR efficiency of 90-110%, and a dynamic range of 8 orders of magnitude.	Pass
qPCR DNA Contamination (E. coli Genomic) A minimum of 1 µl of Luna® Probe One-Step RT-qPCR 4X Mix with UDG (No ROX) is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.	Pass
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of Luna® Probe One-Step RT-qPCR 4X Mix with UDG (No ROX) is incubated at 37°C. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass
Non-Specific DNase Activity (16 hour, Buffer) A 50 µl reaction in 1X Luna® Probe One-Step RT-qPCR Mix with UDG (No ROX) containing	Pass

Assay Name/Specification	Lot # 10176923
1 µg of T3 or T7 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Lea Antonopoulos
Production Scientist
26 Jan 2023



Michael Tonello
Packaging Quality Control Inspector
27 Jan 2023