

New England Biolabs Certificate of Analysis

Product Name: pTYB21 Vector
Catalog Number: N6709S
Concentration: 200 µg/ml
Packaging Lot Number: 10195926
Expiration Date: 05/2026
Storage Temperature: -20°C
Storage Conditions: 10 mM Tris-HCl, 1 mM EDTA, (pH 8.0 @ 25°C)
Specification Version: PS-N6709S v1.0

pTYB21 Vector Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N6709SVIAL	pTYB21 Vector	10191801	Pass

Assay Name/Specification	Lot # 10195926
A260/A280 Assay The ratio of UV absorption of pTYB21 Vector at 260 and 280 nm is between 1.8 and 2.0.	Pass
DNA Concentration (A260) The concentration of pTYB21 Vector is between 200 and 220 µg/ml as determined by UV absorption at 260 nm.	Pass
Restriction Digest (Linearization) A 50 µl reaction in CutSmart® Buffer containing 2 µg of pTYB21 Vector and 10 units of SapI incubated for 1 hour at 37°C produces > 95% linearization resulting in a single band of approximately 7514 bp as determined by agarose gel electrophoresis.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Emily Chen

Emily Chen
Production Scientist
26 May 2023

Michael Tonello

Michael Tonello
Packaging Quality Control Inspector
04 Aug 2023