

## New England Biolabs Certificate of Analysis

**Product Name:** PstI  
**Catalog Number:** R0140M  
**Concentration:** 100,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.  
**Lot Number:** 10036321  
**Expiration Date:** 02/2021  
**Storage Temperature:** -20°C  
**Storage Conditions:** 250 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 0.15% Triton X-100, 200 µg/ml BSA  
**Specification Version:** PS-R0140T/M v1.0

PstI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0140M VIAL	PstI	10036322	Pass
B7203S VIAL	NEBuffer™ 3.1	10021113	Pass
B7024S VIAL	Gel Loading Dye, Purple (6X)	10021138	Pass


Assay Name/Specification	Lot # 10036321
<b>Blue-White Screening (Terminal Integrity)</b> A sample of pUC19 vector linearized with a 10-fold excess of PstI, religated and transformed into an E. coli strain expressing the LacZ beta fragment gene results in <1% white colonies.	Pass
<b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 200 units of PstI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
<b>Ligation and Recutting (Terminal Integrity)</b> After a 100-fold over-digestion of Lambda DNA with PstI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with PstI.	Pass
<b>Non-Specific DNase Activity (16 Hour)</b> A 50 µl reaction in NEBuffer 3.1 containing 1 µg of Lambda DNA and a minimum of 100 units of PstI incubated for 16 hours at 37°C results in a DNA pattern free of	Pass

Assay Name/Specification	Lot # 10036321
<p>detectable nuclease degradation as determined by agarose gel electrophoresis.</p> <p><b>Protein Purity Assay (SDS-PAGE)</b> PstI is &gt;95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.</p>	<p><b>Pass</b></p>

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso  
Production Scientist  
26 Nov 2018



Michael Tonello  
Packaging Quality Control Inspector  
19 Feb 2019