

## New England Biolabs Certificate of Analysis

**Product Name:** *BclI*  
**Catalog Number:** *R0160S*  
**Concentration:** *10,000 U/ml*  
**Unit Definition:** *One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA (dam-) in 1 hour at 50°C in a total reaction volume of 50 µl.*  
**Packaging Lot Number:** *10147625*  
**Expiration Date:** *04/2024*  
**Storage Temperature:** *-20°C*  
**Storage Conditions:** *50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA*  
**Specification Version:** *PS-R0160S/L v1.0*

BclI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0160SVIAL	BclI	10147624	Pass
B6003SVIAL	NEBuffer™ r3.1	10146823	Pass

Assay Name/Specification	Lot # 10147625
<p><b>Non-Specific DNase Activity (16 Hour)</b>            A 50 µl reaction in NEBuffer 3.1 containing 1 µg of Lambda dam- DNA and a minimum of 30 units of BclI incubated for 16 hours at 50°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	Pass
<p><b>Exonuclease Activity (Radioactivity Release)</b>            A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [<sup>3</sup>H] E. coli DNA and a minimum of 100 units of BclI incubated for 4 hours at 50°C releases &lt;0.1% of the total radioactivity.</p>	Pass
<p><b>Ligation and Recutting (Terminal Integrity)</b>            After a 20-fold over-digestion of Lambda dam- DNA with BclI, &gt;95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, &gt;95% can be recut with BclI.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Penghua Zhang  
Production Scientist  
27 Apr 2022



Erin Varney  
Packaging Quality Control Inspector  
27 Apr 2022