

New England Biolabs Certificate of Analysis

Product Name: NruI
Catalog Number: R0192L
Concentration: 10,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Lot Number: 10042311
Expiration Date: 11/2020
Storage Temperature: -20°C
Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R0192S/L v1.0

NruI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0192LVIAL	NruI	10028873	Pass
B7203SVIAL	NEBuffer™ 3.1	10041000	Pass

Assay Name/Specification	Lot # 10042311
<p>Endonuclease Activity (Nicking) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of supercoiled pUC19 DNA and a minimum of 100 units of NruI incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.</p>	Pass
<p>Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 50 units of NruI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.</p>	Pass
<p>Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with NruI, ~50% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with NruI.</p>	Pass
<p>Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of Lambda DNA and a minimum of 100 units of NruI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	Pass

Assay Name/Specification	Lot # 10042311
Protein Purity Assay (SDS-PAGE) Nrul is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass

This product has been tested and shown to be in compliance with all specifications.



Jianying Luo
Production Scientist
21 Nov 2018



Michael Tonello
Packaging Quality Control Inspector
01 May 2019