


## New England Biolabs Certificate of Analysis

**Product Name:** *FspEI*  
**Catalog Number:** *R0662S*  
**Concentration:** *5,000 U/ml*  
**Unit Definition:** *One unit is defined as the amount of enzyme required to digest 1 µg of pBR322 (dcm+) DNA in 1 hour at 37°C in a total reaction volume of 50 µl.*  
**Lot Number:** *10038404*  
**Expiration Date:** *02/2021*  
**Storage Temperature:** *-20°C*  
**Storage Conditions:** *300 mM NaCl , 10 mM Tris-HCl , 1 mM DTT , 0.1 mM EDTA , 50 % Glycerol , 500 µg/ml BSA, (pH 7.4 @ 25°C)*  
**Specification Version:** *PS-R0662S/L v2.0*

FspEI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S0538SVIAL	Enzyme Activator Solution	10036144	Pass
R0662SVIAL	FspEI	10038405	Pass
B7204SVIAL	CutSmart® Buffer	10031567	Pass

Assay Name/Specification	Lot # 10038404
<p><b>Exonuclease Activity (Radioactivity Release)</b>            A 50 µl reaction in CutSmart® Buffer containing 1 µg of a mixture of single and double-stranded [<sup>3</sup>H] E. coli DNA and a minimum of 15 units of FspEI incubated for 4 hours at 37°C releases &lt;0.1% of the total radioactivity.</p>	Pass
<p><b>Non-Specific DNase Activity (16 hour)</b>            A 50 µl reaction in CutSmart® Buffer containing 1 µg of pBR322 DNA and a minimum of 5 of FspEI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: although no nuclease degradation is detected under these conditions, extended incubations and/or high concentrations of this enzyme may result in star activity. See the product FAQ for recommended reaction conditions for this enzyme.</p>	Pass
<p><b>Protein Purity Assay (SDS-PAGE)</b>            FspEI is &gt;95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.



---

Doreen Duquette  
Production Scientist  
06 Feb 2019



---

Michael Tonello  
Packaging Quality Control Inspector  
13 Mar 2019