

New England Biolabs Certificate of Analysis


Product Name: *FspEI*
Catalog Number: *R0662S*
Concentration: *5,000 U/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to digest 1 µg of pBR322 (dcm+) DNA in 1 hour at 37°C in a total reaction volume of 50 µl.*
Packaging Lot Number: *10108701*
Expiration Date: *03/2023*
Storage Temperature: *-20°C*
Storage Conditions: *300 mM NaCl, 10 mM Tris-HCl, 1 mM DTT, 0.1 mM EDTA, 50 % Glycerol, 500 µg/ml BSA, (pH 7.4 @ 25°C)*
Specification Version: *PS-R0662S/L v2.0*

FspEI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S0538SVIAL	Enzyme Activator Solution	10100423	Pass
R0662SVIAL	FspEI	10100422	Pass
B6004SVIAL	rCutSmart™ Buffer	10107576	Pass

Assay Name/Specification	Lot # 10108701
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart® Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 15 units of FspEI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Non-Specific DNase Activity (16 hour) A 50 µl reaction in CutSmart® Buffer containing 1 µg of pBR322 DNA and a minimum of 5 of FspEI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: although no nuclease degradation is detected under these conditions, extended incubations and/or high concentrations of this enzyme may result in star activity. See the product FAQ for recommended reaction conditions for this enzyme.	Pass
Protein Purity Assay (SDS-PAGE) FspEI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Mala Samaranayake
Production Scientist
24 Jun 2021



Michael Tonello
Packaging Quality Control Inspector
24 Jun 2021