

New England Biolabs Certificate of Analysis

Product Name: PI-SceI
Catalog Number: R0696S
Concentration: 5,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to cleave 1 µg of pBSvdeX XmnI-linearized Control Plasmid in 3 hours at 37°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10154978
Expiration Date: 08/2023
Storage Temperature: -20°C
Storage Conditions: 300 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 500 µg/ml BSA
Specification Version: PS-R0696S/L v2.0

PI-SceI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0696SVIAL	PI-SceI	10117694	Pass
N0422SVIAL	pBSvdeX XmnI-linearized Control Plasmid	10154979	Pass
B9200SVIAL	Recombinant Albumin, Molecular Biology G	10150376	Pass
B0696SVIAL	NEBuffer™ PI-SceI	10117697	Pass

Assay Name/Specification	Lot # 10154978
<p>Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of pBSvdeX-XmnI DNA with PI-SceI, ~75% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, ~75% can be recut with PI-SceI.</p>	Pass
<p>Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer PI-SceI containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 50 units of PI-SceI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.</p>	Pass
<p>Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer PI-SceI containing 1 µg of pBSvdeX-XmnI DNA and a minimum of 50 Units of PI-SceI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Penghua Zhang
Production Scientist
18 Jul 2022



Erin Varney
Packaging Quality Control Inspector
18 Jul 2022