

## New England Biolabs Certificate of Analysis

**Product Name:** SNAP-Capture Pull-Down Resin  
**Catalog Number:** S9144S  
**Lot Number:** 10033121  
**Expiration Date:** 01/2021  
**Storage Temperature:** 4°C  
**Specification Version:** PS-S9144S v1.0

### SNAP-Capture Pull-Down Resin Component List

NEB Part Number	Component Description	Lot Number	Individual QC Result
S9144SVIAL	SNAP-Capture Pull-Down Resin	10033314	<b>Pass</b>

Assay Name/Specification	Lot # 10033121
<b>Binding Capacity (Magnetic Beads)</b> SNAP-Capture Pull-Down Resin ( 20 µl ) were equilibrated and incubated with 200 µl of crude extract from E. coli containing a plasmid that expresses a SNAP-intein-chitin binding domain fusion protein for 1 hour at 25°C, then washed and the SNAP CBD fusion cleaved and eluted. Binding capacity was determined to be ≥1 mg of target protein per ml of bed volume.	<b>Pass</b>
<b>Functional Binding Assay (Qualitative)</b> SNAP-Capture Pull-Down Resin ( 20 µl ) were equilibrated and incubated with 200 µl of crude extract from E. coli containing a plasmid that expresses a SNAP-intein-chitin binding domain fusion protein for 1 hour at 25°C, then washed, eluted and evaluated by Tris-Glycine gel to confirm low non-specific binding of extract proteins and high isolation of target.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.



---

Brad Landgraf  
Production Scientist  
08 Jan 2019



---

Michael Tonello  
Packaging Quality Control Inspector  
14 Jan 2019